

UK Declaration of Conformity

BMT MESSTECHNIK GmbH herewith declares under sole responsibility the conformity of the products

OZONE ANALYZER BMT 965

OZONE ANALYZER BMT 964

OZONE MONITOR BMT 932

and all their modifications with the following regulations:

Electromagnetic Compatibility Regulations S.I. 2016 No. 1091

Electrical Equipment (Safety) Regulations S.I. 2016 No. 1101

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012

These products contain less than 5 mg of Hg in their lamps (exemption in accordance to Hazardous Substances and Packaging (Legislative Functions and Amendment) (EU Exit) Regulations 2020, Table 1, No 5: Mercury in other low-pressure discharge lamps).

Harmonized standards applied:

Safety: EN 61010-1: 2010
Safety requirements for electrical equipment for measurement, control, and laboratory use

EMC: EN 61326-1:2013
Electrical equipment for measurement, control and laboratory use – EMC requirements

Serial number: Each instrument has an identifiable serial number.

Year of construction: Each instrument has an identifiable year of construction.

Stahnsdorf, July 19, 2021



Dr. Franz Wallner, President


BMT MESSTECHNIK GMBH
Hamburger Str. 19
14532 Stahnsdorf
+49-3329-69677-0